



ALL RIGHTS RESERVED. NO PORTION OF THIS PUBLICATION, WHETHER IN WHOLE OR IN PART CAN BE REPRODUCED WITHOUT THE EXPRESS WRITTEN CONSENT OF SPC TECHNOLOGY.

SPC-F005.DWG

REVISIONS

DOC. NO. SPC-F005 * Effective: 7/8/02 * DCP No: 1398

DCP #	REV	DESCRIPTION	DRAWN	DATE	CHECKD	DATE	APPRVD	DATE
1430	A	RELEASED	JWM	10/25/04	SF	10/25/04	JC	10/25/04



SPECIFICATIONS

- Test Frequency: 50KHz typical
- Test Range: 1µF or larger
- Power: Two "AA" batteries; 100 hrs for zinc battery and 200 hours for alkaline.

DISCLAIMER:
ALL STATEMENTS AND TECHNICAL INFORMATION CONTAINED HEREIN ARE BASED UPON INFORMATION AND/OR TESTS WE BELIEVE TO BE ACCURATE AND RELIABLE. SINCE CONDITIONS OF USE ARE BEYOND OUR CONTROL, THE USER SHALL DETERMINE THE SUITABILITY OF THE PRODUCT FOR THE INTENDED USE AND ASSUME ALL RISK AND LIABILITY WHATSOEVER IN CONNECTION THEREWITH.

TOLERANCES:
UNLESS OTHERWISE SPECIFIED, DIMENSIONS ARE FOR REFERENCE PURPOSES ONLY.

DRAWN BY:	DATE:
Jeff McVicker	10/25/04
CHECKED BY:	DATE:
Steve Feiwell	10/25/04
APPROVED BY:	DATE:
John Cole	10/25/04

DRAWING TITLE:				ESR Meter			
SIZE	DWG. NO.	ELECTRONIC FILE		REV			
A	72-6355	50N1365.dwg		A			
SCALE: NTS		U.O.M.: INCHES [mm]		SHEET: 1 OF 1			